Fieldpiece

Professional-grade instruments for HVAC/R field service.



This certifies that the following designated products:

LT17A Digital Multimeter

Meets in full the Safety requirements specified on the European Community Directives 73/23/EEC and 93/68/EEC.

Conformity is declared with reference to the following Norms:

IEC/EN 61010-1:2001

Class II, Pollution degree 2

Measuring Terminal: CATIII, 600V, Battery Supply: Single 9V battery

Meets in full the Electromagnetic Compatibility requirements specified on European Community Directives 89/336/EEC.

Conformity is declared with reference to the following Norms:

EN 61326: 2001

1. Emission: EN 55011 class B

2. Immunity: EN 61000-4-2 (2001); EN 61000-4-3 (2003); EN 61000-4-4 (2002)

Ryan Cunningham U C

ı ∪ 03/23/2022

Mechanical Engineer, QA